

## RIMBA GLOVE SDN BHD

### PRODUCT SPECIFICATION

EFFECTIVE DATE: 29/07/2021

REV NO: 01

**Product Information:**

Trade Name	Rimba Glove - Powder Free Nitrile Examination Gloves - Non Sterile
Intended Use	To conduct medical examination, diagnostic and therapeutic procedures to protect patient and user from cross contamination of infection.
Product Conformance	ASTM D6319 / EN 455 series / EN 420:2003 +A1:2009, EN ISO 374-1:2016, EN 374-2:2014, EN 16523-1:2015, EN 374-4:2013 and EN ISO 374-5: 2016
Material	Carboxylated Acrylonitrile - Butadiene Synthetic Rubber Latex
Weight	3.5gm
Color	Blue
Type	Ambidextrous
Surface finish	Half-Finger Textured

**Product Specification**

DIMENSION	Size	Palm Width (mm)	Length (mm)
	X-Small	70 ± 10	min 240
	Small	80 ± 10	min 240
	Medium	95 ± 10	min 240
	Large	110 ± 10	min 240
	Extra Large	120 ± 10	min 240

  

Thickness	Single wall thickness (mm)		
	Cuff	min 0.05	
	Palm	min 0.07	
	Finger	min 0.12	

  

Physical Properties		Un-aged	Aged
	Tensile Strength(Mpa)	min 14	min 14
	Ultimate Elongation (%)	min 500	min 450
	Force at break (N)	min 6 N	min 6 N

  

Quality Specification	Inspection Criteria	Sampling Plan/AQL
	Dimension	S2 - AQL 4.0
	Physical Properties/Force at Break	S2 - AQL 4.0
	Water Tight Test	G1-AQL 1.5
	Visual Inspection - Major	G1-AQL 2.5
	Visual Inspection - Minor	G1-AQL 4.0
	Residual Powder	n = 5 (average) < 2 mg/g
	Packaging labeling	G1- AQL 1.5
	Glove Physical Counting	S2 - AQL 1.5 ± 2%

**Packaging Configuration and Labeling**

Packing	100 gloves per inner dispenser (By weight) 10 inner per outer carton
Labeling	PER PACKING SPECIFICATION
Lot Numbering System	Per Rimba Glove Internal Lot Number / Customer assigned Lot Number